



13th September 2023

NITRILE DISPOSAL GLOVES FOOD GRADE DECLARATION

In relation to the following products:

- iKON Nitrile gloves blue color long cuff, M5.8g, length 300mm

PRODUCT CODE	PRODUCT DESCRIPTION
IKGLVXLNITLC	IKON NITRILE GLOVE L/CUFF 1XL 100P/1000C
IKGLVLNITLC	IKON NITRILE GLOVE L/CUFF LRG 100P/1000C
IKGLVMNITLC	IKON NITRILE GLOVE L/CUFF MED 100P/1000C
IKGLVSNITLC	IKON NITRILE GLOVE L/CUFF SML 100P/1000C

We hereby confirm that the above mentioned products supplied by DUNNINGHAMS are FDA approved and comply with the following regulation:

- Gloves are also manufactured using food grade raw materials that are approved for direct food contact and conform to the following standards:

- HACCP International Food Safety Certification System
- EU Regulation EN 1186 - Materials and articles in contact with foodstuffs
- FDA Title 21 CFR (Code Federal regulation) 177.2600 rubber articles intended for repeated use.

- Other Purposes for Use: The Nitrile Examination Gloves are also a non-sterile disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner.

Regulation Number: 21 CFR 880.6250

Regulation Name: Non-Powdered Patient Examination Glove

Regulatory Class: Class I, reserved

Gloves are to be stored in a clean dry area, out of direct sunlight, where temperatures don't exceed 30°C and not below - 5°C.

VALIDITY DATE: This declaration is valid until the next relevant legislative and/or regulatory change.

Regards

Philippa Avis
Quality Assurance Manager
DM Dunningham Ltd